# Surface Temperature Sensor (Order Code STS-BTA)



The Surface Temperature Sensor is designed for use in situations in which low thermal mass or flexibility is required. Special features include an exposed thermistor that results in an extremely rapid response time.

Typical uses for the Surface Temperature Sensor include the following:

- Skin temperature measurements
- Human respiration studies
- Specific heat experiments
- Heat transfer experiments
- Friction and energy studies

**Note:** Vernier products are designed for educational use. Our products are not designed nor are they recommended for any industrial, medical, or commercial process such as life support, patient diagnosis, control of a manufacturing process, or industrial testing of any kind.

## What's Included

• Vernier Surface Temperature Sensor

## **Compatible Software and Interfaces**

See www.vernier.com/manuals/sts-bta for a list of interfaces and software compatible with the Surface Temperature Sensor.

# **Getting Started**

- 1. Connect the sensor to the interface (LabQuest Mini, LabQuest 2, etc.).
- 2. Start the appropriate data-collection software (Logger *Pro*, Logger Lite, LabQuest App) if not already running, and choose New from File menu.

The software will identify the sensor and load a default data-collection setup. You are now ready to continue your experiment.

If you are collecting data using a Chromebook<sup>TM</sup>, mobile device such as iPad<sup>®</sup> or Android<sup>TM</sup> tablet, or a Vernier wireless sensor or interface, please see the following link for up-to-date connection information:

#### www.vernier.com/start/sts-bta

#### **Using the Product**

Connect the sensor following the steps in the Getting Started section of this user manual.

The Surface Temperature Sensor is for use in air and water only. For temperature measurements in harsher environments that require a more durable probe, we recommend our Stainless Steel Temperature Probe (order code TMP-BTA).

# Videos

View videos related to this product at www.vernier.com/sts-bta

# Calibration

It should not be necessary to perform a calibration when using this sensor. It is calibrated before it ships. However, you can calibrate the sensor using Logger *Pro* 3.3. or newer. **Note:** Calibration can only be done on computers running Logger *Pro*. For more information, see www.vernier.com/til/1310

## **Specifications**

Temperature range	-25 to 125°C (-13 to 275°F)
Maximum temperature that the sensor can tolerate without damage	150°C
13-bit resolution	0.04°C (-25 to 0°C)
	0.02°C (0 to 40°C)
	0.05°C (40 to 100°C)
	0.13°C (100 to 125°C)
	0.08°C (-25 to 0°C)
12-bit resolution	0.08°C (-25 to 0°C)
	0.03°C (0 to 40°C)
	0.1°C (40 to 100°C)
	0.25°C (100 to 125°C)
10-bit resolution	0.32°C (-25 to 0°C)
	0.12°C (0 to 40°C)
	0.4°C (40 to 100°C)
	1.0°C (100 to 125°C)
Temperature sensor	20 k $\Omega$ NTC Thermistor
Accuracy	±0.2°C at 0°C, ±0.5°C at 100°C
Response time (time for 90% change in reading)	50 seconds (in still air)
	20 seconds (in moving air)
Probe dimensions: probe length (handle plus body)	15.5 cm

# How the Sensor Works

This probe uses the 20 k $\Omega$  NTC Thermistor, which is a variable resistor. When the temperature increases, the resistance decreases non-linearly. The best-fit approximation to this nonlinear characteristic is the Steinhart-Hart equation. At 25°C, the resistance is approximately 4.3% per °C. The interface measures the resistance value, *R*, at a particular temperature and converts the resistance using the

Steinhart-Hart equation:

## $T = [K_0 + K_1(\ln 1000R) + K_2(\ln 1000R)^3]^{-1} - 273.15$

where *T* is temperature (°C), *R* is the measured resistance in k $\Omega$ ,  $K_o = 1.02119 \times 10^{-3}$ ,  $K_I = 2.22468 \times 10^{-4}$ , and  $K_2 = 1.33342 \times 10^{-7}$ . Our programs perform this conversion and provide readings in °C (or other units, if you load a different calibration).

# Troubleshooting

First check that the sensor is responding: Connect the sensor and start the data-collection program. Hold the tip of the sensor in your hand to warm it and check that the temperature changes.

For additional troubleshooting and FAQs, see www.vernier.com/til/1391

# **Repair Information**

If you have watched the related product video(s), followed the troubleshooting steps, and are still having trouble with your Surface Temperature Sensor, contact Vernier Technical Support at support@vernier.com or call 888-837-6437. Support specialists will work with you to determine if the unit needs to be sent in for repair. At that time, a Return Merchandise Authorization (RMA) number will be issued and instructions will be communicated on how to return the unit for repair.

# Warranty

Vernier warrants this product to be free from defects in materials and workmanship for a period of five years from the date of shipment to the customer. This warranty does not cover damage to the product caused by abuse or improper use. This warranty covers educational institutions only.



#### MEASURE. ANALYZE. LEARN."

Vernier Software & Technology 13979 SW Millikan Way • Beaverton, OR 97005-2886 Toll Free (888) 837-6437 • (503) 277-2299 • Fax (503) 277-2440 info@vernier.com • www.vernier.com

Rev. 3/18/16

Logger *Pro*, Logger Lite, Vernier LabQuest 2, LabQuest Mini, and other marks shown are our trademarks or registered trademarks in the United States.

All other marks not owned by us that appear herein are the property of their respective owners, who may or may not be affiliated with, connected to, or sponsored by us.